# User Manual

revision e
**ucm 2**

Wheelchair Master Remote for the DX BUS System
*with*
built-in Wireless Computer mouse
*and*
optional built-in Wireless Mobile Phone interface
*and*
Environmental Control-Communication Aid interface.

Important Notes:

1. Read this manual carefully before installing or operating your Genie+.
2. Due to continuos product improvement Unique Perspectives reserves the right to update this Manual. This Manual supersedes all previous issues which must not continue to be used.
3. Any attempt to gain access to or in any way abuse the electronic components of the Genie+ renders the manufacturer's warranty void and the Manufacturer free from liability.
4. Parts of this document have been reproduced with kind permission of Controls Dynamics Ltd. from their 'DX Dolphin Remote Installation Manual' dated Jan 1999.

# Contents

This manual is divided into two sections. Section 1 is aimed for Users, Carers, Occupational Therapists and Technicians to quickly and easily understand the basic requirements for operating the Genie+ control safely. Section 2 contains further information for the Technician and Occupational Therapist to understand the installation, adjustment and fault finding procedures.

In both sections the Maintenance and Safety and Misuse chapters are presented. This information in particular must be read and understood before operating the Genie+.

## Section 1 – User Operation


| | | |
|---|---|---|
| 1 | **Introduction** | 6 |
| 2 | **Features** | 8 |
| 3 | **Operation** | 12 |
| | Turning on the Genie+ | 12 |
| | Driving the chair | 12 |
| | Selecting another operation mode | 12 |
| | Speed | 14 |
| | Horn | 15 |
| | Seat function | 16 |
| | Lights | 17 |
| | Computer access | 17 |
| | Mobile phone | 18 |
| | External Access | 18 |
| | Power Off | 18 |
| | Making an Alarm | |
| 4 | **TypeMatic On-Screen Keyboard software** | 20 |
| | Installing the software | 20 |
| | Running the program | 20 |
| | Editing the layout | 22 |
| 5 | **Batteries & Charging** | 28 |
| | Battery type | 28 |
| | Battery charging | 28 |
| | Battery gauge | 28 |
| | Battery saver | 28 |
| 6 | **Maintenance** | 28 |

7 **Safety and Misuse Warnings** 30
Know the Risks and Limitations 31

## Section 2 – Installation, Adjustment & Fault Finding

1 **Related Documentation** 33

2 **Specifications** 34
Genie+ 34
Mouse Emulator 36

3 **Installation and Testing** 37
Mounting 37
Connections 37
Switch inputs 38
Switch outputs & interfacing to other devices 39
Interfacing to a USB port 41
DX programming socket 43
DX BUS cabling 43
USB Cable 44
Testing 45
Power Up Response 45
Genie+ check sequence 45
Speed check sequence 46
Horn check sequence 46
Seat function check sequence 46
Lights check sequence 46
Computer access check sequence 47
Mobile phone check sequence 47
External Access check sequence 48
Power Off check sequence

4 **Programming the Genie+** 50
Programmable Parameters 50
Selecting program mode 57
LED meaning in program mode 57
Parameter Adjustment Table – Page One 59
Parameter Adjustment Table – Page Two 60
Parameter Adjustment Table – Page Three 61
Creating Your own Set of Operation Modes 62
Resetting Parameters to Factory Values 62

| | | |
|---|---|---|
| 5 | **Programming the DX System** | 64 |
| | Downloading the correct DX wizard file | 64 |
| | Programming Methods & Tools | 67 |
| | Dynamic Wizard | 67 |
| | Hand Held Programmer (HHP) | 67 |
| | Auto download | 68 |
| 6 | **Battery Warnings, Diagnostics and Fault Finding** | 72 |
| | Battery warning conditions | 72 |
| | Battery high warning condition | 72 |
| | Battery Low warning condition | 72 |
| | Low capacity warning condition | 72 |
| | Diagnostics and Fault Finding | 73 |
| | Flash Code | 73 |
| | Limp Mode | 76 |
| 7 | **Maintenance** | 77 |
| 8 | **Safety and Misuse Warnings** | 78 |
| | Know the Risks and Limitations | 79 |
| 9 | **EC Declaration Of Conformity** | 80 |
| 10 | **Warranty** | 82 |
| 11 | **Sales and Service Information** | 83 |
| 12 | **Genie+ Parameter Quick Reference** | 84 |
| | **History of improvements and changes** | 85 |

# SECTION ONE

# USER OPERATION

# 1 Introduction

The Genie+ is a master remote display for the DX bus system. It enables integrated control over a wheelchair, computer, mobile phone and environmental or communication device from any DX or 3rd party remote joystick module (RJM). Available RJM's include:

1. Joystick Module *DX-RJM*
2. Heavy Duty Joystick *DX-RJM-HD*
3. Finger Steering Control *DX-RJM-VIC*
4. Four Switch and Five Switch Interface *DX-ACC4, DX-5SW*
5. Sip and Puff Module *DX-SNP*
6. Mini Joystick *DX-RJM-MINI*
7. HeadFirst head array

Just one switch or a flick of the joystick is all that is required to switch modes from, for example, driving the wheelchair to moving the computer mouse and back again. It is simple to use with no complex menus or LCD displays.

The Genie+ can be fitted to any wheelchair that is fitted with a DX power module and a DX remote joystick module (The Genie+ is a master remote and therefore replaces the original joystick).

InfraRed transmitter for mouse emulator and BlueTooth module for phone are located behind front panel.
System Status LED
7-segment LED displays driving speed (drive profile).
Battery gauge
Program button
Genie+
On/Off and Menu switch. Can also use external switch.
External switch input
DX BUS socket
Interface to external device (ECU, AAC)

On the font of the unit is the Genie+ switch. The Genie+ switch is used to turn on the chair, and, in combination with the RJM, to select one of eight functions; Speed, Horn, Seat function, Lights, Computer Mouse, Mobile phone control, External device control (environmental control or communication aid) and finally Power off.

If the user is unable to operate the Genie+ switch on the unit itself an external switch can be used. If the user is unable to use switches at all a special “joystick flick” mode is available whereby a backwards flick of the remote joystick takes the place of switch selections.




In the centre of the display a 7 segment display indicates the driving speed or drive profile. Around this display is a pattern of 8 LEDs which indicate the operating mode. When none of these LEDs are lighting the chair is in drive mode. At the top left hand side of the display is the battery gauge. The DX system status LED is located at the top right hand side.

The Genie+ has a standard DXBUS connector so that it may be connected to the DX power module. The Genie+ can be used to control any powered wheelchair that is fitted with a DX Power Module and a DX Remote Joystick module.

The Genie+ and the associated DX Power Module are fully programmable to cater for a wide range of chair types and user needs. Correct installation and programming are essential to ensure optimum performance and safety.

The operation mode of the Genie+ can be programmed without the need of a DX Hand Held Programmer or a Dynamic Wizard. Programming of the Genie+ is performed on the device itself.

## Example Genie+ System



# 2 Features

**Built-In Wireless Computer Mouse.**

The Genie+ has a wireless computer mouse built-in. This is an infra-red link to the Genie+ mouse emulator which connects into the USB port of your computer. When computer access is selected the joystick on the Genie+ is used to move the mouse on the screen. Pressing the Genie+ switch activates the left click.

**Mobile Phone Interface**

The Genie+ can be supplied with an internal bluetooth module that enables control over a SmartPhone. The joystick can be used to browse the contents of your phone's contact list, make and receive calls, and send text messages.

**Alarm**

The Genie+ can be enabled to emit an audible alarm and send a text message when the Genie+ switch is held for a defined period of time. To send a text message alarm you must have a Genie+ equipped with a mobile phone interface and the SmartPhone must be switched on and connected.

**On-Screen Keyboard included.**

The Genie+ mouse emulator is supplied with a free copy of TypeMatic on-screen keyboard software. This software can be used to type text into any Windows application and includes powerful word prediction and completion features.

**Mode selection with just one switch**

The Genie+ accommodates people who can operate a joystick but who find it difficult to operate the buttons on a standard joystick control. With the Genie+ just one switch or a flick of the joystick is all that is required to change operating modes.

**Driving Mode**

Standard wheelchair driving mode. Drive profile is displayed in the 7 segment display.

**Seat function Mode**

A seat symbol is displayed in the 7 segment display. Left/Right movements of the joystick selects which seat actuator to operate. Forward/Back movements of the joystick operate the chosen actuator. The 7 segment display provides feedback to the user as to which actuator is selected.

| | |
|---|---|
| **Lights Mode** | A forward movement of the joystick turns on or off the head lights. A backwards movement of the joystick turns on or off the hazard lights. Left/Right movements of the joystick turn on or off the indicators (unless the hazards are on). |
| **Computer Access** | A “C” is displayed in the 7 segment display. By moving the joystick the mouse on the computer is moved about the screen. The Genie+ switch operates the left click. To exit Computer Access mode the Genie+ switch is held for a predefined length of time. |
| **External Access** | The Genie+ can be connected to an environmental control or communication aid device. The joystick can then be used to make selections on the external device. For example the joystick can be used to highlight a cell on a grid of communication symbols using left/right and up/down movements. The cell can be selected by clicking the Genie+ switch.<br><br>To exit External Access mode the Genie+ switch is held for a predefined length of time. |
| **Direct connection to GEWA PROG III environmental control transmitter.** | The Genie+ can be interfaced to any environmental control transmitter but is designed to be specifically connected to a GEWA PROG III. In this instance the button on the PROG is highlighted using left/right and up/down movements of the joystick and selected by pressing the Genie+ switch. |
| **USB Interface** | A USB interface cable allows the Genie+ to be connected to the USB port of an “on-chair” computer or communication device. When external access is selected the joystick on the Genie+ is used to move the mouse pointer on the external device. The Genie+ switch operates the Left Click or Select. |
| **Drive Profile Select** | When selected with a left deflection the drive profile speed is changed. |
| **Power Off** | This is a special mode that simply turns off the chair. |
| **Mode Selection** | When the Genie+ switch is pressed the Genie+ enters mode selection. In this mode Forward/Back deflections of the joystick are used to highlight each “row” or “pair” of functions. For example Speed and Horn, Seat and |

Lights etc. Left/Right deflections are used to select the desired function in the “pair”, i.e. the left or right function. So for example to choose the Lights mode the user must press the Genie+ switch, make a back deflection and then a right deflection.

For users who are unable to press the Genie+ switch or an external switch a special “flick” mode is available whereby a backwards flick of the joystick takes the place of pressing the switch.

For users who are used to the mode selection methods of the Genie+ joystick product these metheds are still available to Genie+ users. How to select them is described further on in the manual.

**Programming button**

The selection method of the Genie+, the number and type of operating modes and other options can be programmed without the use of a computer or other external device. The program button is located on the front panel and pressing this for over 10 seconds sets the Genie+ in program mode. Pressing it for less than 10 seconds switches off the Genie+. All the options can be set by using the Genie+ switch, the program switch the 7 segment display and the battery gauge.

**RJM compatible**

The Genie+ is compatible with any DX or 3rd party remote joystick module. These include:

1. Joystick Module
2. Heavy Duty Joystick
3. Finger Steering Control
4. Four Switch and Five Switch Interface
5. Sip and Puff Module
6. Mini Joystick
7. HeadFirst head array

**Battery gauge**

Battery charge level is indicated by a set of 6 LEDs.

**System status**

Faults within the DX system are indicated by flashing the system status LED.

**Programming socket HHP / Wizard socket**

The standard HHP / Wizard socket for programming the DX system. Located underneath the device.

**Standard DXBUS connection**

For connecting the Genie+, with a DXBUS cable to the DX power module.

**Battery charger**

Due to space restrictions the battery charger socket is not located on the Genie+ itself but on a separate DX BUS lead (DX part no: DX-ACC3).

**3.5mm Switch Inputs**

For people who cannot use the Genie+ switch you can connect external switches such as TASH buddy buttons etc.

# 3 Operation

### Turning on the Genie+.

To turn on the Genie+ press The Genie+ switch. The 7 segment display will show a "-" symbol and after a short moment will display the current drive profile.

If you have the "Accidental_Hits" option enabled you must hold Switch1 for at least 1 second before the Genie+ will power up properly. This avoids involuntary movements on an external switch.

For more information on Genie+ option see Section 2 Chapter 4.

### Driving the Chair.

Driving is the default operation mode. It is automatically selected when the wheelchair is turned on or when another operation mode is finished.

To drive the chair forward deflect the joystick forward. To drive the chair backwards deflect the joystick backwards. To drive the chair to the right deflect the joystick to the right. To drive the chair to the left deflect the joystick to the left.

**Warning:** If the driving performance is poor or erratic consult your service agent immediately.

### Selecting another operation mode.

There are 8 other operation modes apart from driving. These are:-

| Mode | Symbol |
|---|---|
| 1. Speed | |
| 2. Horn | |
| 3. Seat Function | |
| 4. Lights | |

| | | |
|---|---|---|
| 5. | Computer | |
| 6. | Mobile Phone | |
| 7. | External device |  |
| 8. | Power Off | |

**Method 1 – Switch + Joystick (Default method)**

1. Press the Genie+ switch

2. The 7 segment display will show a "-" sign and the top "row" or "pair" of symbols will illuminate, namely the Speed and Horn symbols.

3. Use Forward and Back deflections of the joystick to select the "row" or "pair" of symbols which contains the mode you require.

4. Once you have the correct "row" or "pair" selected use a Left or Right deflection of the joystick to select the mode you require.

5. The Genie+ enters that mode.

**NOTE:** If you wish to exit the selection mode and go back to driving without choosing an operating mode press the Genie+ switch.

**Method 2 – Joystick Only**

1. Make a backwards "flick" of the joystick.

2. The 7 segment display will show a "-" sign and the top "row" or "pair" of symbols will illuminate, namely the Speed and Horn symbols.

3. Use Forward and Back deflections of the joystick to select the "row" or "pair" of symbols which contains the mode you require.

4. Once you have the correct "row" or "pair" selected use a Left or Right deflection of the joystick to select the mode you require.

5. The Genie+ enters that mode.

**Other Methods**

For compatibility with the Genie+ Joystick there are 4 additional methods of menu selection. These are identical to those found in the Genie+ Joystick product. This allows a user who is migrating from a Genie+ Joystick to a Genie+ to use their new device in a way that they are already familiar with.

---

**WARNING:** Once a mode is being selected or when a mode has been entered it is no longer possible to drive the wheelchair. It is extremely important, therefore, to be in a safe location before selecting or entering another operation mode. Driving is only possible again after the mode has been finished.

In method 1 to go back to driving a user must press the Genie+ switch (pressing and holding the switch in computer mouse, mobile phone and external access modes).

In method 2 to go back to driving the user makes a backwards “flick” of the joystick.

---

**Speed (Drive Profile)**

When Speed is selected (with a left deflection of the joystick) the 7 segment display indicates the current drive profile number. If the joystick is held deflected the speed begins to change. Release the joystick when the desired speed number is displayed.

Note that the number of Drive profiles that can be selected is limited by the Max_Profile_Number. For details on changing this and other options see Section 2 Chapter 4.

**Horn**

When Horn is selected (with a right deflection of the joystick) the Genie+ emits a loud beep.

## Seat Function

When seat function is selected the Genie+ enters seat function mode and the joystick is used to select, extend and retract actuators (so long as they are fitted to the wheelchair).

Left/Right deflections of the joystick select 1 of the 5 actuators to control. A flashing segment in the display indicates which actuator is currently selected as the table below illustrates.

| Back Rest | Back Tilt | Seat Height | Foot Rest L | Foot Rest R |
|---|---|---|---|---|
| | | | | |
| Actuator 1 | Actuator 2 | Actuator 5 | Actuator 3 | Actuator 4 |

Forward/Back deflections of the joystick are used to extend or retract the selected actuator.

**NOTE:** It is possible to edit the actuator list so that only those actuators that are fitted on the chair are displayed in seat function mode. See Section 2 Chapter 4.

To exit seat function mode press and release the Genie+ switch.

When the selection method is set to “Joystick Only” exit seat function mode with a backward “flick” of the joystick.

## Lights

When lights is selected the Genie+ enters Lights mode and the joystick is used to turn on and off the wheelchair lights.

? A forward deflection of the joystick turns on/off the sidelights.

? A backwards deflection of the joystick turns on/off the hazard lights.

? Left/Right deflections of the joystick turn on/off the indicators.

The seven segment display indicates the status of the wheelchair lights according to the following table:-

| Side Lights | Left Indicator | Right Indicator | Hazards |
|---|---|---|---|
| |  | | |

To exit lights mode press and release The Genie+ switch.

When the selection method is set to “Joystick Only” exit lights mode with a backward “flick” of the joystick.

## Computer Mouse

When computer mouse is selected the Genie+ enters computer mouse mode and the joystick is used to control the mouse on the computer (provided that the mouse emulator is connected and the wheelchair is within range. See page 35 for details on fitting the mouse emulator).

A forward deflection of the joystick moves the mouse pointer up the screen. A backward deflection of the joystick moves the mouse pointer down the screen. A left deflection of the joystick moves the mouse pointer to the left and a right deflection of the joystick moves the mouse pointer to the right.

Pressing the Genie+ switch performs a Left Click.

In Windows XP it is possible to change the speed of the mouse pointer. See page 47.

To exit computer mouse mode press and hold the Genie+ switch until the Genie+ emits a beep. The length of time that the switch must be held can be adjusted by the “Exit_Time” parameter.

For people who are unable to use the computer keyboard an on-screen keyboard software entitled “TypeMatic” is provided. This allows a user to select keys and words from a grid on the screen and type into any Windows program. The software is discussed in the next chapter.

Joystick only mode

If the selection method is set to “Joystick Only” the computer mouse mode operates in a different way.

A left “flick” of the joystick performs a left click. A right “flick” of the joystick performs a right click. A forward “flick” of the joystick performs a “double left click” and a backwards “flick” of the joystick exits computer mouse mode.

Note that in this mode the mouse will not begin to move until after the “flick” time, which is set by the “Time_Base” parameter.

**Mobile Phone (optional)**

When mobile phone is selected the Genei+ enters mobile phone mode and the joystick is used to control a Smart Phone. The phone must be Bluetooth equipped and running Windows Mobile 2003 operating system. Joystick deflections are used to navigate the menus on the phone and make and receive telephone calls. On screen keyboard software running on the phone allows you to compose and send text messages. For further details on using your Genie+ to control a Smart Phone and how to install the software see the “ClickToPhone” user manual.

To exit mobile phone mode press and hold the Genie+ switch until the Genie+ emits a beep. The length of time that the switch must be held can be adjusted by the “Exit_Time” parameter.

Joystick Only selection mode

When the selection method is set to “Joystick Only”…

? a forward “flick” of the joystick is the equivalent of pressing the Genie+ switch.

exit mobile phone mode with a backward “flick” of the joystick.



**External Access.**

When external access is selected the Genie+ enters external access mode and the joystick is used to control the external device (provided that the external device has been connected to the Genie+ with the correct cable. See page 38 for details on interfacing to external devices.

Devices with a 5 switch input

These devices can be controlled in exactly the same way as if a non-proportional joystick, such as a TASH mini joystick or Star switch, were directly connected to it. The Genie+ takes the place of such a switch.

The Genie+ expects that the device presents a grid of options to the user and that forward/back and left/right deflections will highlight an individual cell in the grid whilst pressing the Genie+ switch will select and activate the highlighted cell. The grid might be a set of communication symbols on a communication aid (for example a Cameleon CV) or the buttons on an environmental control transmitter (for example the GEWA prog. See section 2 Chapter 4 on how to interface with a GEWA Progress).

Devices with 2x3.5mm switch inputs
These devices would normally be controlled by 2 switches and would implement some sort of scanning mechanism. One switch is used to scan the cells of a grid whilst the other is used to activate the highlighted cell.

With the Genie+ the switches are replaced by deflections of the joystick. A forward deflection is used to scan the cells of the grid whilst a backwards deflection is used to activate the highlighted cell. Again the grid could be a set of communication symbols on a communication aid (a Words+ MessageMate for example) or the buttons of an environmental control transmitter (a TASH Relax for example).

Controlling 2 external devices
If 2 external devices are connected they must both be controlled as described above except that Left/Right deflections of the joystick are used for the second device.

To exit external access mode press and hold The Genie+ switch until the Genie+ emits a beep. The length of time that the switch must be held can be adjusted by the “Exit_Time” parameter.

Joystick Only selection mode
When the selection method is set to “Joystick Only”…

- a forward “flick” of the joystick is the equivalent of pressing The Genie+ switch.
- exit external access mode with a backward “flick” of the joystick.

**Power Off**

When power off is selected the Genie+ turns off.

You can also turn off the wheelchair by pressing the program button for less than 10 seconds. It is not intended that a user would use this method.

## Making an Alarm.

In any mode the user can generate an alarm by pressing and holding The Genie+ switch until an ‘A’ is displayed in the 7 segment display and the Genie+ starts beeping rapidly. There are 2 types of alarm.

1- Audible alarm
2- Audible alarm + Text Message Alarm

The type of alarm is set by the “Alarm” parameter.

When the second type of alarm is chosen and the Genie+ is equipped with a BlueTooth interface and the SmartPhone is switched on and connected the Genie+ instructs the SmartPhone to send a pre-written text message is sent to a specific number. For more information see the “ClickToPhone” user manual.

The length of time that the switch must be held to generate an alarm can be adjusted by the “Exit_Time” parameter.

For details on changing this and other options see Section 2 Chapter 4.

# 4 Type-Matic Software

## Installing the Software

1. Insert the CDROM into your CD-ROM drive
2. If the installation program does not “autorun” then choose ‘Run’ from the Windows Start menu and continue with step 3, otherwise jump to step 5.
3. Type d:\setup.exe. (if your CD-ROM drive is not drive D, type the appropriate letter instead.)
4. Choose OK.
5. Follow the instructions on the screen.

## Running the Program

Run TypeMatic by choosing Programs from the Start menu and clicking on the TypeMatic icon.

When you run TypeMatic for the first time the following default Grid is presented:-


Standard keyboard keys.
Type-Matic
1. a b c d e f g h i j @ #
2. k l m n o p q r s t . !?
3. u v w x y z F
5 Special keys. Spacebar, Backspace, Enter, Shift and SentenceBar. To view help text on a character position the mouse over the cell and pop-up help text will appear.
Links to other grids for numbers, punctuation and function keys.

Now launch a word processing program such as WordPad and open a new document. To type, position the mouse pointer over the desired character and click. The character should appear in your new document.

Typematic's word prediction displays a list of words beginning with the characters you have typed. In the example below the 'h' character has been typed.

To choose one of the predictions simply click on it and watch it appear in your word document.

As you type TypeMatic will remember any new words you create.

When you type a space, TypeMatic tries to predict the next word in the sentence you are typing. Because this prediction is based on previous sentences you have written this prediction will not begin right away. The more you repeat certain sentences the more TypeMatic will learn to predict the next word in the sentence.

TypeMatic's abbreviation file, 'shorthand.txt', contains a list of abbreviations. Initially it contains only one abbreviation, 'hau'. Type 'hau', then enter and watch 'How are you' appear in your document. You can edit shorthand.txt located in the application directory and create your own abbreviations. Be careful to only use character groupings which do not constitute a word in themselves.

To quit the program right click on the window and choose 'Exit' from the Pop-Up menu. If the 'UpdateFile' property is set to 'Prompt' you will be asked whether or not you want to save any new words you have written. Choose Yes or No as required.

### Setting the Dwell time

For users who are unable to use Left or Right click buttons a Dwell feature is provided whereby a cell is automatically selected by keeping the mouse pointer

still for a certain length of time over the desired cell. This time is called the dwell time and can be set between 1 and 5 seconds. Right click on any cell in the grid and select Dwell time from the PopUp menu.

Now when you position the mouse pointer over a cell and keep it still for dwell time the contents of the cell will be typed into the active application, i.e. the cell is automatically selected without having to make a click.

Furthermore when you select a Dwell time a set of buttons representing Left click, Left Drag and Right click appear in the left hand side of the TypeMatic window. These allow you to generate mouse clicks in other applications by keeping the mouse pointer still for the dwell time over the control you want to click.




First select the mouse click you want to generate by keeping the mouse pointer still over the button representing the desired mouse click. After the button is selected (down position) move the mouse pointer over the control you want to click in the other application and keep it still for the dwell time.

### Using the Sentence Bar

The Sentence Bar feature allows a user you to pre-prepare a sentence or word before sending it to the active application. To activate the sentence bar Right Click on any cell in the grid and click on Sentence Bar in the pop-up menu.

The Sentence bar will appear at the top of the TypeMatic window.

When you have finished typing your word or sentence click on the Enter cell to send the text in the sentence bar to the active application.

To de-activate the sentence bar Right Click on any cell in the Grid and click on Sentence Bar in the pop-up menu.

The Sentence bar can also be activated from a cell in the Grid identified by the following icon:

You would use this method if a person is unable to perform a Right Click.

**Note:** If this cell does not appear in the Grid you are using, but you require it, then you must edit the Grid and create a cell with the “Key” property set to “SENTENCE_BAR”. This is one of the special keys selected from the drop down list. See the following page on how to edit the cells in the Grid.

In Windows 2000 and Windows XP it is not possible to input text letter by letter into certain text boxes. Examples of these text boxes include the FileName text box in WordPad® and the URL text box in Internet Explorer®. You will know when this happens because as you type each letter it overwrites any text that was previously in the text box. The result is that you can only type one letter. To overcome this limitation use the Sentence Bar.

**Note:** When you close TypeMatic it will remember if the Sentence Bar was activated or not. If you always use the Sentence Bar you will only have to activate it once.

## Editing the Grid

To edit a Grid right click on any cell and choose ‘Edit Grid…’ from the pop-up menu. The Layout properties window is presented.




The left hand column displays the property names. The right hand column displays the property values. The panel at the bottom of the window displays the properties help text.

To view the help text of a property click on the property name. The help text for the property appears in the panel at the bottom of the window.

When you click on the property name a control appears in the property value field. The type of control depends on the property type, for example a drop down list for Grid Size, a file open button for Picture, a text entry control for Caption. Use the control to edit the property value.

## Properties

| | |
|---|---|
| Layout | You can have any number of grids with TypeMatic. Click the FileSave button to save the current Grid or click the FileOpen button to open another. |
| BackColor | The color of the background. |
| BevelStyle | The bevel style of cells in the grid. Can be set to none, Inset or Raised. |
| BevelWidth | The width of a cell bevel. Can be set between 1 and 8. |
| CellGap | The gap between cells. Can be set between 0 and 8. |
| CellSize | The size of the cells in the grid. Can be set to 16x16, 32x32 or 64x64. |
| Font | The font used throughout the grid. |
| GridPosition | The start-up position of the grid on the screen. |
| GridSize | The size of the grid. The smallest size is 4x4 cells, the largest is 8x10 cells. |
| HighlightColor | The color used to highlight cells. |
| Abbreviations | Specifies whether or not the abbreviation feature is enabled. The abbreviation list is stored in the file 'shorthand.txt' in the application directory. You can edit it to add more abbreviations using notepad. |
| WordPrediction | The type of word prediction used. Can be set to none, current word or current plus next word. When set to none the word prediction column is hidden. |
| UpdateFile | Specifies how the word prediction file 'typematic.dic' is updated with the words written and chosen by a user. Can be set to Always, Never or Prompt. If set to prompt the user is asked whether or not they wish to save any new words when they quit the program. |
| *CellColor | The color of a cell. If the cell is transparent this property has no effect. |
| *FontColor | The color of the font in the cell. |
| *Key | The key(s) to be pressed (simulated) when the cell is selected. Use '+' for SHIFT, '%' for ALT and '^' for CTRL. These special keys have a toggle function. To specify an actual plus sign use '{+}'. For '^' use '{^}', for '%' use '{%}'. You can select other special keys from the listbox.<br><br>Note than when you select a key the text is automatically assigned to the caption. You can have the caption text different from the key text but always set the key text first, then the caption text. |
| *Caption | The text that appears in the cell. |
| *HelpText | The help text for the cell which will be displayed when the mouse is paused over the cell. |

| | |
|---|---|
| *Transparent | Specifies whether or not a cell is transparent. A transparent cell takes on the color of the background when not selected and the color of the highlight when selected. |
| *Picture | Each cell can contain a picture (bitmap or icon only). |
| *NextGrid | You can link a cell to another grid by specifying the file here. In this way you can have links to grids of numbers and punctuation. Always remember to have a link back to the main grid! |

### Editing the properties of a cell

Properties beginning with an '*' are specific to an individual cell and only effect the selected cell. To edit the property of a cell, for example it's key, click on the cell in the main window, then click the property name called 'key' and edit it's content.

A special case is the column of word predictions. These cells have the same properties. In other-words if you set the cell color for a word prediction cell it effects all word prediction cells.

### Editing the 'general' properties of the Grid

Properties which do not begin with a '*' effect the general look and functionality of the Grid. For example the GridSize is a general property.

### Saving and Opening layout files

When you have finished editing properties you can save the Grid as a file on your hard disk. The files are called 'tlf' files: 'Typematic Layout File'. To save a Grid click on the first property called 'layout'. A File Open and File Save button appear in the property value field. Click the File Save button to save your newly edited Grid.

Typematic comes with a selection of files to illustrate different Grids. These files are called small.tlf, standard.tlf and colorful.tlf and are located in the application directory. To load a Grid click on the first property called 'layout' then click on the File Open button, locate the desired file and choose OK.

**Note:** When you close the Layout properties window TypeMatic will remember which Grid is in use and will load this Grid the next time it is run.

You can also open grids directly from the main TypeMatic window by right clicking on any cell and choosing Open Grid from the popup menu.

## Editing the Dictionary

Included with the TypeMatic installation is a utility program called MakeDictionary.exe which can be found in the program directory. This program allows you to edit the dictionary or create a new one from a text file.




**Tip:** If you wish to merge a text file with the existing dictionary, for example to add a vocabulary set to a user's dictionary, first open the dictionary file and then open the text file. You will be prompted as to whether you wish to discard the exiting words. Choose No. Once the new words are loaded choose Make Dictionary.

# 5 Batteries and Charging

## Battery Type

The DX System is designed to perform optimally with either Lead-Acid or Gel Cell 24 V deep cycle batteries, rated at 20 – 120 Amp hours. The maximum average discharge rate must not exceed half the rated capacity, in Amp hours.

## Battery Charging

Due to space restrictions the battery charger socket is not located on the Genie+ itself but on a separate DX BUS lead (DX part no: DX-ACC3)..

Progress of the charge can be monitored by turning on the Genie+ and watching the battery gauge.

## Battery Gauge

The battery gauge provides true, useable battery capacity information. A full battery with at least 85% of rated capacity, is represented by all 6 LEDs lit. Some new batteries can start with as little as 80% capacity, developing higher capacity in their early life (sometimes up to 110%), before slowly deteriorating over their rated life.

As the battery voltage drops, the number of LEDs lit reduces from top to bottom. When only the red LEDs are lit, the available battery capacity is typically less than 10%. At this level and below, the Battery Gauge flashes to alert the user that the wheelchair is running on reserve capacity. The battery capacity will reduce more rapidly in the reverse capacity range.

## Battery Saver

The Battery Saver is a feature programmed into the DX Remote. When the battery capacity is in the reserve range (below 21V), the wheelchair performance is reduced. This is to preserve the life of the battery by encouraging the user to recharge the battery before it becomes harmfully flat.

Operating the wheelchair with more than two LEDs of the battery gauge lit will generally give normal wheelchair performance. This is provided that the battery size and the PM program settings are matched to the wheelchair.

# 6 Maintenance

1. The Genie+ system should be regularly checked for integrity. Loose, damaged or corroded connectors or terminals, or damaged cabling should be reported to your Service Centre and be replaced immediately.

2. The cabling of Genie+ System including Switch cables, DXBUS cable, Charger cable and Battery cables, should be regularly checked for integrity. They should never be loose. Cables should be neatly attached to the wheelchair frame and mounts so that no possibility exists for a cable to become snagged on the moving parts of the wheelchair itself, the person sitting in the wheelchair, and/or items external to the wheelchair such as door handles etc.

3. All switches connected to the Genie+ should be regularly tested to ensure that they function correctly.

4. During storage and transport of your wheelchair ensure that there is no possibility that the Genie+ switch can inadvertently by pressed thereby causing the chair to turn on and possible enter a drive state. Always disengage the motor gears and disconnect any external switches.

5. Under no condition should a latching switch be connected to the Genie+. Only connect non-latching switches.

6. The Genie+ components and other wheelchair parts should be kept free of dust, dirt and liquids. If necessary wipe with a cloth dampened with warm water or alcohol. **Do not** use solvents or abrasive cleaners.

7. Where any doubt exists, consult your nearest Service Centre or Agent.

8. There are no user-serviceable parts within the Genie+. Do not attempt to open the case.

9. In accordance with the requirements of CE marking of this device and the Company's policy, it is requested that re-occurring faults or defects are reported back to Unique Perspectives Ltd.

**Warning !!** If the Genie+ is damaged in any way, or if internal damage may have occurred (for example by being dropped), have it checked by qualified personnel before operating.

# 7 Safety and Misuse Warnings

Do not install, maintain or operate this equipment without reading, understanding and following the proper instructions and manuals, otherwise injury or damage may result.

The completed installation must be thoroughly checked, and all programmable options must be correctly adjusted for safe operation prior to use.

A warning must be conveyed to the wheelchair operator that the controller could cause the chair to come to a sudden stop. In situations where this may affect the safety of the user, this will require the fitting and wearing of a seat belt.

Performance adjustments should only be made by professionals of the health care field or persons fully conversant with this process and the driver's capabilities. Incorrect settings could cause injury to the driver, bystanders, damage to the chair and surrounding property.

After the wheelchair has been set up, check to make sure that the wheelchair performs to the specifications entered in the programming procedures. If the wheelchair does not perform to specifications, turn the wheelchair off immediately and re-program. Repeat procedure until the wheelchair performs to the specifications.

Do not operate the DX system if it behaves erratically, or shows abnormal response, heating, smoke or arcing. Turn the system off, disconnect the battery or open the battery overload switch, and consult your service agent.

Do not operate your DX system if the battery is nearly flat as a dangerous situation may result due to a loss of power in an inopportune place.

Ensure the controller is turned off when not in use.

No connector pins should be touched, as contamination or damage due to electrostatic discharge may result. Dummy sockets in unused DXBUS connectors should be left in place unless a new module is added to the system.

Whilst designed to resist water penetration, under certain conditions moisture might enter the Genie+. Any spillage's over the Genie+ should be wiped dry without delay. The Genie+ may be used outdoors in light drizzle conditions but should be protected from rain.

Most electronic equipment is influenced by Radio Frequency Interference (RFI). Caution should be exercised with regard to the use of portable communications equipment in the area around such equipment. While the manufacturer has

made every effort to ensure that RFI does not cause problems, very strong signals could still cause a problem. If RFI causes erratic behavior, shut the wheelchair off immediately. Leave off while transmission is in progress.

In the event of a fault indicator flashing while driving (battery gauge and/or status LED), the user must ensure that the system is behaving normally. If not, the system must be turned off and a Service Agent called immediately.

Report any malfunctions immediately to your Service Agent.

## Know the risks and limitations

Like any mechanical propelled vehicle there are certain risks involved.

The driver is responsible for any damage or injury that may occur to a party as a result of using a powered wheelchair. If the driver cannot assume responsibility due to age or disability then a carer must be present and be able to take over control either using a stop switch or a dual control in case of an emergency. You may wish to consider taking out insurance to cover any claims arising from such an incident.

The most sensitive part of a Genie+ system is the joystick element itself. The owner or carer must assume responsibility for regularly checking the integrity of the joystick element and report any problems to the service agent immediately.

# SECTION TWO

# INSTALLATION, ADJUSTMENT & FAULT FINDING

# 1 Related Documentation

A DX based wheelchair control system may comprise between two and sixteen DX compatible modules depending on the application. Each DX compatible module has its own User Manual which describes the installation requirements of that particular module.

This manual describes the Genie+ remote only and must therefore be read in conjunction with the:

- Relevant DX RJM Installation Manual
- DX Power Module (PMB) Installation Manual
- DX Hand Held Programmer (HHP) Manual
- Dynamic Wizard Installation Sheet
- Installation Manuals for all other DX Modules to be used in your application.

# 2 Specifications

## Genie+

### Electrical

| | |
|---|---|
| Compatible with standard DXBUS | |
| Operating voltage range | 18v – 32v d.c. |
| Quiescent Current | <1mA Off, typically 250mA On |

### Mechanical

| | |
|---|---|
| Weight | Approx 0.5 Kg |
| Mounting | As required by installer |
| Case material | Extruded aluminum, plastic coated. |
| Size | 85mmx105mmx35mm |

### Environmental

| | Min | Max | Units |
|---|---|---|---|
| Operating ambient temperature range | -25 | 50 | ºC |
| Storage temperature range | -25 | 70 | ºC |
| Operating and storage humidity | 0 | 90 | %RH |

The Genie+ is designed to resist water penetration, but under certain conditions moisture might enter the control. Suitable for light precipitation.

### Intended Use

The Genie+ is designed to be used as a Master Remote Display for the Controls Dynamic DX Control System for Powered Wheelchairs.

The Genie+ contains a DX UCM (User Control Module) and is therefore a Master Remote.

**Warning !** In any DX System there can be only one master remote connected at any one time. The system will not function correctly if more than one Master Remote are connected.

The Genie+ is compatible with the following DX power modules:

- DX-PM, DX-PM1, DX-PM2

The Genie+ is compatible with the following Actuator modules:

- DX-CLAM, DX-TAM, DX-ARC5

The Genie+ is intended to be used in conjunction with a DX Secondary remote. The Genie+ is compatible with the following secondary modules:

- *DX-RJM*
- *DX-RJM-HD*
- *DX-RJM-VIC*
- *DX-ACC4, DX-5SW*
- *DX-SNP*
- *DX-RJM-MIN*

For compatibility with any other DX Module introduced by Controls Dynamic please contact Unique Perspectives.

The Genie+ can be operated with an external switche rather than the button on the unit itself if required. The external switche should be of high quality, professionally connected and regularly maintained. For details of wiring and connections see Chapter 3 of this section.

The Genie+ is intended for use on Class A and B powered wheelchairs only.

The Genie+ is not designed for use with any other control system.

# Mouse Emulator

## Electrical

| | |
|---|---|
| Power Supply | 5V 100ma DC from USB interface |
| Quiescent Current | 100mA |

## Mechanical

| | |
|---|---|
| Weight | Approx. 250grams |
| Case material | Extruded aluminum, painted black. |
| Size | 115mmx55mmx35mm |

## Environmental

| | Min | Max | Units |
|---|---|---|---|
| Operating ambient temperature range | -25 | 50 | ºC |
| Storage temperature range | -25 | 70 | ºC |
| Operating and storage humidity | 0 | 90 | %RH |

The Mouse Emulator is not designed for outdoor use.

## Intended Use

The Mouse Emulator is a USB IBM compatible pointing device designed to enable those individuals who cannot use a standard computer mouse the ability to manipulate the windows mouse pointer using the Genie+ in computer access mode.

The Mouse Emulator receives proprietary IrDA InfraRed signals from the Genie+ and converts them into USB signals.

The Mouse Emulator is for indoor use only.

The Mouse Emulator can only be connected to a PC with a USB port. A PS/2 version is available for older computers on request.

The Mouse Emulator can only be used with a Genie+.

# 3 Installation & Testing

## Mounting

A mounting system for the Genie+ is available from Unique Perspectives. Contact your supplier for details. This consists of a mounting plate and swing away arm. The mounting plate is fitted to the Genie+ with 2xM4 screws.

**Warning !!** The installation of the mount should be carried out by a qualified technician or other professional. For safe fitting of the mounting plate to the Genie+, select a screw length between 6mm and 10mm.

Dimensions of the mounting positions on the underside of the unit are shown below:



# Connections

The connection panel is located at the bottom of the device and is illustrated below.




## Switch Inputs

For users who cannot operate the button on the Genie+ itself a 3.5mm jack socket is provided on the connection panel. You can connect any non-latching switch to this socket provided it is fitted with 3.5mm jack plugs.

| Connector | Pin | Signal |
|---|---|---|
| Standard 3.5mm jack plug | Tip | Switch |
| | Sleeve | Switch common |

**Warning!!**

A user's switches must be connected professionally to the Genie+ by a qualified technician. Unique Perspectives accept no responsibility or liability for poorly made connections which may result in incorrect operation and possibly dangerous operation of the Genie+.

Under no condition should a latching switch be connected to the Genie+. Only use non-latching switches.

Unused sockets should be fitted with the plugs provided.

## Switch Outputs & Interfacing to other devices.

To connect the Genie+ to an external device such as a communication aid or environmental control device five solid state relay contacts are provided on a 6 pin Mini DIN connector. These contacts reflect the state of the RJM joystick position and the Genie+ switch when external access mode has been selected by the user.

| 6 pin Mini DIN. Connector | Pin | Signal |
|---|---|---|
|  | 1 | Switch common |
| | 2 | Joystick Left |
| | 3 | Joystick Right |
| | 4 | Joystick Down |
| | 5 | Joystick Up |
| | 6 | The Genie+ switch |



- *relays are current rated to 250ma*
- *contacts are isolated from wheelchair electronics*

The interface to the external device will depend upon the number of switch inputs and connector type available on the external device. Several different interface options are described below:-

**Devices with a 9 pin D switch input**

If the external device is capable of being controlled by a five switch device such as a TASH mini joystick, wafer pad or star switch then device will have a 9 pin D switch input. With a 6 pin Mini DIN to 9 pin D cable the Genie+ can be connected to such devices and operated in exactly the same way as if the five switch device were connected to it. The pinout of the 9pin D is shown below:

| Connector | Pin | Signal |
|---|---|---|
| 9 Pin Male D  Male D Front View | 1 | Forward (Switch 2) |
| | 2 | Backwards (Switch 3) |
| | 3 | Left (Switch 4) |
| | 4 | Right (Switch 5) |
| | 5 | |
| | 6 | Select (The Genie+ switch) |
| | 7 | |
| | 8 | Switch common |
| | 9 | |

A 6 pin Mini DIN to 9 pin D cable is available for connection to devices with this type of switch input.

**Devices with 3.5mm jack switch inputs**

Unfortunately most communication aid devices have only 2x3.5mm switch inputs. In this case the Genie+ can be wired to the device so that a forward deflection of the joystick activates the the Genie+ switch input of the external device (normally the "select") and a backwards deflection of the joystick activates the switch 2 input of the external device (normally the "advance" or "step").

A 6 pin Mini DIN to 2x3.5mm jack plug cable is available for connection to such devices.

See USB method on following page as an alternative interface method for communication devices

**The GEWA PROG environmental control interface**

The switch input of the GEWA PROG III is also a 6 pin Mini DIN connector whose pin-out matches that of the Genie+. To operate the PROG from the joystick it is necessary to put the PROG in joystick input mode. You can do this by pressing "**P+5"**, followed by "**3"** on the PROG. Consult the PROG manual for further details.

A pin-to-pin Mini DIN cable is available for connection to this device

**Interfacing to GEWA Progress ECU**.

A special interface cable is required whose pin-out is described below and the Genie+ parameter "External Device" must be set to "3". See Section 2 Chapter 4 for details on adjusting Genie+ parameters. The Progress should be set for 2 switch manual row/column scanning. When external access is selected, downward deflections of the joystick select the row. Right hand deflections select the cell within that row. And a final downward deflection activates the chosen cell. It is important to carry out the sequence of deflections as described otherwise the Progress may get out of step with the Genie+. If this happens simply exit external access mode and re-enter it.

| Genie+ Pin | Signal | | Progress Pin | Signal |
|---|---|---|---|---|
| 1 | Switch common | → | Jack 1&2, Sleeve | Switch common |
| 2 | Joystick Left | → | | |
| 3 | Joystick Right | → | Jack 2, Tip | Switch 2 signal |
| 4 | Joystick Down | → | Jack 1, Tip | The Genie+ switch |

| | | | | |
|---|---|---|---|---|
| | | | | signal |
| 5 | Joystick Up | → | | |
| 6 | The Genie+ switch | → | | |

A Mini DIN cable to 2 3.5mm jacks is available for connection to this device

**Interfacing to an environmental control unit and a communication aid**.
To interface to more than 2 external devices it is necessary to operate both of them in a 1 or 2 switch mode as per “Devices with 3.5mm jack switch inputs” described above. Use forward/back deflections of the joystick for one device and left/right deflections for the other. An interface cable for this kind of setup is available on request. It is important to note however that as most modern communication aids have a built in environmental control unit, or the option of it, this type of setup is rarely required.

**Interfacing to DynaVox communication aid with switch inputs**.
The switch inputs of a DynaVox are peculiar in that they can not share a common signal. A special interface cable is required whose pin-out is described below and the Genie+ parameter “External Device” must be set to “2”. See Section 2 Chapter 4 for details on adjusting Genie+ parameters.

| Genie+ Pin | Signal | | DynaVox Pin | Signal |
|---|---|---|---|---|
| 1 | Switch common | → | | |
| 2 | Joystick Left | → | Jack 1, Sleeve | The Genie+ switch common |
| 3 | Joystick Right | → | Jack 2, Sleeve | Switch 2 common |
| 4 | Joystick Down | → | Jack 2, Tip | Switch 2 signal |
| 5 | Joystick Up | → | Jack 1, Tip | The Genie+ switch signal |
| 6 | The Genie+ switch | → | | |

A Mini DIN cable to 2 3.5mm jacks is available for connection to this device. Note that for modern devices with a USB port the USB interface is a better solution, see below.

**Interfacing to the USB port on an “on-chair” computer or communication device**

A USB interface cable is now available. This allows the Genie+ to be connected to the USB port of an “on-chair” computer or communication device. The cable connects from the Genie+ switch output port to the USB port of the external device. The cable has electronics built in which “fool” the computer or communication device into thinking that a standard USB mouse is connected.

When external access is selected the joystick on the Genie+ is used to move the mouse pointer and the Genie+ switch makes left click selections. Mouse movements are non-proportional.

When interfacing to a communication device such as a DynaVox Mk4, MiniMo or MightMo all that is required is to set the Input Method to Mouse. When interfacing to a computer no set up is required. The computer sees the Genie+ as an ordinary USB mouse.

Note that this interface is in addition to the wireless mouse already built into the Genie+. This means that the user can have wireless control over a desktop computer, and also control (through the cable) of an “on-chair” computer.

**Warning!!** Unique Perspectives has tested the external access function with a number of external devices including products from Cambridge Adaptive Communication, GEWA AB, Toby-Churchill and the DynaVox range. Unique Perspectives cannot guarantee correct operation of other external devices. The relay contacts are solid state and current rated to 250ma. Although the relay connections are isolated extreme care must be taken when wiring to new products. Unique Perspectives accept no responsibility or liability for poorly made or incorrect connections which may result in incorrect operation and possibly dangerous operation of the Genie+ and/or external device.

## DX Programming Socket

The DX hand held programmer or DX Wizard cable plugs into this socket which is located underneath the unit. This gives the programmer access to editing the parameters of the DX System to optimize wheelchair performance and drive characteristics.

## DX BUS Cabling

The Genie+ has one DXBUS connector which enables the Genie+ to be connected to the DX system.

The Genie+ will normally be connected directly to one of the two Power Module DXBUS sockets with a DXBUS cable.

The Charger cable plugs into the second DXBUS socket.

If you are connecting more DX modules such as a CLAM or RJM you will need to fit a DX BUS 4 Way Socket to expand the number of available sockets. The part no. for this socket is DX-SKT-X4.

The DXBUS cables are available in the following standard lengths

DXBUS Cable, Straight, 0.12M
DXBUS Cable, Straight, 0.3M
DXBUS Cable, Straight, 0.5M
DXBUS Cable, Straight, 1.0M
DXBUS Cable, Straight, 1.5M
The DXBUS cable length should be selected so that the cable is neatly attached to the wheelchair frame. The cable should never be loose and no possibility should exist for the cable to snag on the moving parts of the wheelchair, the person sitting in the wheelchair, and/or on items external to the wheelchair such as door handles etc.

A typical DXBUS connection system is illustrated on the following page:




## USB Cable

The Mouse Emulator is fitted with a 2m USB cable.

The USB Cable is a standard cable that is connected to the USB port of your computer. However a PS2 version is available on request for older computers.

# Testing

Ensure that all DX Modules used in the system and the Genie+ have been installed as specified in their installation procedures. The Genie+ needs to be correctly programmed for the appropriate wheelchair prior to testing. This is normally done by the supplier, see chapters 4 & 5 of this section.

A Genie+ Remote contains the complete wheelchair system set up, from which all attached modules download their relevant information when the DX system is first turned on.

## Powering Up

Power up the Genie+ by pressing and releasing the Genie+ switch.

## Power Up Response

The power up response for the Genie+ is:

- The 7 segment display will display the current Drive Profile.
- The System Status LED will come on steady.

Note: The first time the Genie+ is turned on after it has been fitted, the System Status LED will flash a fault. This is because the Genie+ must download its information to the DX Power Module. Turn the Genie+ off, then on, to clear this fault.

- At least one of the LEDs on the Battery gauge will be on.

## Genie+ Check Sequence

Perform the following Genie+ check sequence:

Note: The following check sequence assumes that the parameters of the Genie+ have been reset to default (as shipped). For more information on resetting parameters to default see page 61.

1. Turn on the joystick by pressing The Genie+ switch.
2. If the system status LED does not come on steady, but flashes, then there is a fault with one of the DX modules in the system or there has been an auto-download.
3. Confirm that the Genie+ is driving the chair by moving the joystick.

**Speed (Drive profile) check sequence**

1. Press the Genie+ switch and confirm that the Drive Profile number changes when you keep the joystick deflected to the left.

**Horn check sequence**

1. Press the Genie+ switch and confirm that the Horn sounds when you keep the joystick deflected to the right.

**Seat Function check sequence (if actuators are fitted)**

1. Press the Genie+ switch and use the joystick to select Seat function.

1. Use Left/Right deflections of the joystick to select an Actuator.

2. Use Forward/Back deflections of the joystick to extend or retract the actuator.

3. Press the Genie+ switch to return to driving.

**Lights check sequence (if lights are fitted)**

2. Press the Genie+ switch and use the joystick to select Lights.

3. Confirm that the head lights are operating by deflecting the joystick forward.

4. Confirm that the hazard lights are operating by deflecting the joystick backwards. Turn them of by deflecting the joystick a second time.

5. Confirm that the indicators are operating by deflecting the joystick Left/Right.

6. Press the Genie+ switch to return to driving.

## Computer Access check sequence

1. Connect the Genie+ mouse emulator to the USB port of your computer and then turn on the computer.

2. After Windows starts confirm that the green LED behind the front panel of the emulator is blinking. This indicates that the USB connection is established but that the signals from the Genie+ on the wheelchair are not being received.

3. Position the emulator at the edge of the table in line with the computer monitor.

4. Turn on the wheelchair by pressing the Genie+ switch on the Genie+ and position it facing the computer monitor.

5. Confirm that the green LED behind the front panel of the emulator goes steady. This indicates that the infrared link between the Genie+ and the mouse emulator is working correctly.

   *If the green LED does not come on steady then the joystick is out of range, not positioned properly or there is an obstruction in the way. The emulator should be able to pick up the signals from a distance of 2 metres with a ±30° spread in all directions when the Genie+ is mounted on the wheelchair at an angle of 45° (the actual InfraRed transmitter is mounted at a 45° angle within the Genie+). This ability can deteriorate with the effect of ambient lighting. It is important to position the emulator overlapping the edge of the table (or ideally mounted to the underneath of the table) so that the table itself does not obstruct the infrared signals between the two devices.*

6. Press the Genie+ switch and use the joystick to select computer access.

7. Confirm that the RJM joystick is moving the mouse pointer on the screen.

8. Confirm that the Genie+ switch is operating the Left Click.

9. Press and hold the Genie+ switch until the Genie+ emits a beep to return to driving (about 2 seconds).

### Setting the mouse speed

In Windows XP it is possible to adjust the Genie+ mouse speed. Setting the optimum mouse speed for a particular user is an important part of the installation procedure. Too slow and the user will find moving from one side of

the screen to the other time consuming. However too fast and the user will be unable to make the small movements required to select certain controls, for example when browsing a menu. Choosing the best speed is a compromise between speed and controllability.

*Although the default Genie+ mouse speed is Fast, for best control it should be set to Slow and the mouse speed increased in the XP Control Panel. See page 52.*

## To edit the mouse pointer speed

**1.** From the “Start” menu choose “Control Panel” and then click on “Printers and Other Hardware”.




**2.** Click on “Mouse”.




**3.** Click on the “Pointer Options” tab.




**4.** Select the required pointer speed by adjusting the “Motion” slider. Note that changing the “Enhance pointer precision” option also affects mouse speed. We recommend that this option is disabled.




## Mobile Phone check sequence

1. Refer to the ClickToPhone manual.

## External Access check sequence (if fitted)

1. Connect the Genie+ to the external device using the correct interface cable. See page 38 for details.

2. Press the Genie+ switch and use the joystick to select external access.

3. Confirm that the joystick is operating the external device as expected.

4. Press and hold the Genie+ switch until the Genie+ emits a beep to return to driving (about 2 seconds).

**Power Off check sequence**

1. Press the Genie+ switch and use the joystick to select power off.

2. Confirm that the Genie+ turns off.

***And***

1. Turn on the Genie+ again.

2. Press and release the program button and confirm that the Genie+ turns off.

---

**Warning**: If any of the checks fail contact your supplier immediately and do not continue to use the Genie+.

---

# 4 Programming the Genie+

**Warning !!**
Incorrect or inappropriate programming of the Genie+ can put the wheelchair into a dangerous state. Unique Perspectives accept no responsibility or liability for accidents caused by incorrect programming. This section must be read and understood before attempting to program the Genie+

Ensure that the programmed Genie+ complies with all prevailing regulatory requirements for your country and application

## Introduction

The performance of the Genie+, its operating mode and parameter settings may not be decided or known until the chair is supplied to the end user by their assessment center or other supplier. Typically the fine tuning of the Genie+ will be carried out by a technician in collaboration with an occupational therapist during assessment of a users operation capability, very often in the user's home or place of work. It is for this reason that the programming of the Genie+'s performance is carried out on the device itself without the need for a DX hand held programmer or a DX Wizard program.

The Genie+ has a set of programmable parameters which allow a technician in collaboration with an occupational therapist or other professional to adjust the Genie+ so that it best suits a user's needs and switch/joystick operation capability. The programmable parameters are described below.

## Programmable Parameters Page 1

### Selection_Method

This parameter selects the required selection method, i.e. how the user changes modes.

Selection Method = 1, Switch Only.
Selection Method = 2, Switch & Joystick.
Selection Method = 3, Joystick & Switch
Selection Method = 4, Joystick Only.
Selection Method = 5, Genie+ Switch & Joystick (Default)
Selection Method = 6, Genie+ Joystick Only.

Methods 1-4 are provided for compatibility with the Genie+ Joystick product only. It is envisaged that only modes 5 and 6 will be used with the Genie+.

"Switch Only"

By pressing and holding the Genie+ switch the display toggles through symbols representing each mode. Release the Genie+ switch when the desired mode is displayed.

"Switch & Joystick"
Press the Genie+ switch to enter mode selection. The display shows a symbol representing the first mode. Deflecting the joystick forward goes onto the symbol for the next mode. Deflecting the joystick backwards goes onto the symbol or the previous mode. Pressing the Genie+ switch selects the mode that is displayed.

"Joystick & Switch"
Press the Genie+ switch to enter mode selection. The display shows a symbol representing the first mode. Pressing the Genie+ switch again goes on to the symbol for the next mode. Deflecting the joystick forward selects the mode that is displayed.

"Joystick only"
A backwards flick of the joystick selects mode selection. The display shows a symbol representing the first mode. Deflecting the joystick forward goes onto the symbol for the next mode. A backwards flick of the joystick selects the mode that is displayed.

"Genie+ Joystick & Switch"
Press the Genie+ switch to enter mode selection. The first "row" or "pair" of symbols lights up. Use Forward/Back deflections of the joystick to select the "row" or "pair" containing the desired mode. Use Left/Right deflections of the joystick to select the mode.

"Genie+ Joystick only"
A backwards flick of the joystick selects mode selection. The first "row" or "pair" of symbols lights up. Use Forward/Back deflections of the joystick to select the "row" or "pair" containing the desired mode. Use Left/Right deflections of the joystick to select the mode.

## Time_Base

This parameter sets the time base for the various selection methods.

In "Switch Only" it is the length of time that the switch must be held in order to move on to the next mode.

In "Joystick + Switch" it is the repeat rate when the joystick is held forward or backwards.

In "Joystick only" and "Genie+ Joystick only" it is the joystick repeat rate but more importantly the length of time that defines a joystick "flick" is half this time base parameter.

When the accidental hit parameter is set to 2 or 4 the time base defines the switch acceptance time.

The parameter can be set to one of eight values:-

| | |
|---|---|
| 1 = time base of 250ms (¼ second) | 5 = time base of 600ms |
| 2 = time base of 300ms | 6 = time base of 800ms |
| 3 = time base of 350ms | 7 = time base of 1.2seconds |
| 4 = time base of 450ms (Default) | 8 = time base of 2seconds |

## Avoid_Accidental_Hits

1 = none
2 = Mode selection
3 = On power up only
4 = Both

Power Up
When set to 3 or 4 this parameter requires a user to hold the Genie+ switch for 2 seconds before the Genie+ will fully power up.

Mode Selection
When set to 2 or 4 the Genie+ switch must be pressed for a certain amount of time before a switch press is recognised. By setting this option involuntary presses of the switch can be avoided.

The length of time is half the time base. In other-words if the time base is 1.2 seconds the switch 'acceptance time' is 0.6 seconds.

## Switch 2

Unused in the Genie+

## Power_Down_Timer

When set to "Y" the Genie+ will automatically power down after 20 minutes if no activity is detected.

## Exit_Time

The length of time that The Genie+ switch must be held on order to return to driving mode from Computer Access mode, Phone mode or External Access mode. In all other modes pressing the Genie+ switch immediately returns to driving.

1 = 500 ms, 2 = 1 sec, 3 = 2 sec, 4 = 3 sec,
5 = 4 sec, 6 = 5 sec, 7 = 6 sec 8 = 7 sec

It is important to consider whether you are controlling an environmental control device in external access when setting this parameter. If so remember that some functions may require the Genie+ switch to be held for a reasonable length of time, such as a TV volume UP command. It is for this reason that this parameter exists. Set the Exit_Time long in this type of setup. If you are not

connecting an environmental control then it is better to keep the Exit_Time as short as possible in order to return to driving as quickly as possible.

## Programmable Parameters Page 2

### Alarm

This parameter defines what kind of alarm will be made. To activate the alarm the user must press and hold the Genie+ switch until an 'A' is displayed in the 7 segment display and the Genie+ starts beeping rapidly. This length of time is equal to twice the Exit_Time.

1 – No alarm
2 – Audible alarm
3 – Audible alarm + Text Message

When set to 3 and the Genie+ is equipped with a BlueTooth interface and the SmartPhone is switched on and connected the Genie+ instructs the SmartPhone to send a pre-written text message is sent to a specific number. For more information see the "ClickToPhone" user manual.

### Max_Profile_Number

This parameter defines the range of Drive profile numbers that the user can select on the Genie+. Note that the Max Profile Number parameter in the DX wizard file should always be set to 5. See page 63 for programming DX parameters.

### RJM

Must be set to 'Y' to enable the DX remote joystick module.

### Enable_My_Set

The Genie+ offers the unique possibility of creating your own set of operation modes to choose from and in what order you choose them. For example one user may only want to switch between driving and computer access whist another may wish to switch between driving, computer access, seat function and drive profile selection. How to create this set is described later in this chapter.

After creating your set of operation modes you must enable it by setting this parameter to 'Y' and setting the Selection Method parameter to 1,2,3 or 4.

The default set of operation modes and the order in which they are presented is:

1. Computer Access

2. External Access
3. Phone
4. Seat control
5. Lights,
6. Select drive profile,
7. Power down.

## Joystick_Threshold

Unused in the Genie+

## External Device

This parameter selects what type of external device you are connected to. It is specifically to designed to overcome the peculiarities of interfacing to a DynaVox communication device and to provide a more user friendly interface to a GEWA Progress ECU.

1 = standard device, such as GEWA Prog3, CAM3, TASH Relax, Tellus, and any device that can already be controlled using switches.
2 = 2 cable connection to a DynaVox communication device
3 = 2 cable connection to a GEWA Progress

DynaVox communication aids
These devices have 2 switch inputs with isolated common signals. Because the Genie+ uses a shared common on the switch outputs it is necessary to use a special interface cable and to enable this parameter. See page 38 for wiring details on the interface cable.

## Programmable Parameters Page 3

This page allows the seat function list to be edited so that only those actuators that are present on the wheelchair will be displayed. It allows you to associate a symbol with a particular actuator or to not display it. The last parameter in this page determines the base mouse speed for RJM joysticks.

The standard seat function list is illustrated below:

| Parameter | Parameter / Symbol Name | Symbol | Actuator (default) |
|---|---|---|---|
| 1 | Back Rest | | 1 |
| 2 | Back Tilt | | 2 |
| 3 | Seat Height | | 5 |
| 4 | Foot Rest Left | | 3 |
| 5 | Foot Rest Right | | 4 |

### Back_Rest

This parameter selects which Actuator is associated with the symbol. Setting the parameter to 6 means that it will not be displayed in the scan. Default is Actuator 1.

### Back_Tilt

This parameter selects which Actuator is associated with the symbol. Setting the parameter to 6 means that it will not be displayed in the scan. Default is Actuator 2.

### Seat_Height

This parameter selects which Actuator is associated with the symbol. Setting the parameter to 6 means that it will not be displayed in the scan. Default is Actuator 5.

## Foot_Rest_Left

This parameter selects which Actuator is associated with the symbol. Setting the parameter to 6 means that it will not be displayed in the scan. Default is Actuator 3.

## Foot_Rest_Right

This parameter selects which Actuator is associated with the symbol. Setting the parameter to 6 means that it will not be displayed in the scan. Default is Actuator 4.

## RJM_Mouse_Speed

This parameter selects the base mouse speed when using an RJM joystick.

1 = slow (default)
2 = medium
3 = fast

## Programming Mode

All programmable parameters can be adjusted on the Genie+ itself without the need for a DX hand held programmer or a DX Wizard program.

### To select the Programming Mode

1. The programming switch is located on the front panel of the Genie+ and identified by the spanner symbol .
2. Connect the Genie+ to a DX system.
3. Turn on the Genie+ by pressing The Genie+ switch.
4. Press and hold the programming switch for longer than 10 seconds. The Genie+ will make a beep. Release the program switch to enter programming mode.

### LED meaning in program mode.

The parameters are divided into 4 separate groups or ‘pages’.

Page 1 is identified by 1 flash of the decimal point in the 7 segment display. To move on to the next page press the program switch.

Page 2 is identified by 2 flashes of the decimal point. To move on to the next page press the program switch.

Page 3 is identified by 3 flashes of the decimal point. To move on to the next page press the program switch.

Page 4 is identified by 4 flashes of the decimal point.

To exit program mode press the program switch one more time. The decimal point flashes rapidly to indicate the end of program mode and the saving of all parameters after which normal operation continues.

The battery LEDs of the Genie+ are used to indicate which parameter within a particular page is being adjusted. There are six parameters in each page and the current parameter is identified by the number of illuminated battery LEDs. The exception is page 4 which does not use the battery gauge and is for creating your own set of operation modes. This is discussed further on.

The 7-segment display indicates the value of a particular parameter. The parameter can be a number or a ‘Y’ or ‘N’ symbol.


Number of battery LEDs indicates parameter.
7-segment display indicates value of parameter.
Number of flashes of the decimal point indicates the current page.
Genie+

## Selecting a parameter

The Genie+ switch is used to step from one parameter to the next and to change a parameter value. A short click of The Genie+ switch selects the next parameter.

## Changing a parameter value

To change the value of a parameter press and hold The Genie+ switch until the Genie+ emits a beep and the parameter value changes.

***Tip**: Before you enter program mode, look at the table of parameters and know beforehand what changes you want to make. Photocopy the 'Quick Reference Guide' at the end of this manual and by circling the options you have chosen you will have a record of your personalised settings.*

The tables on the following pages list the Genie+ parameters.

## Parameter Adjustment Table – Page One

| Parameter (page 1) | Genie | Value and meaning |
|---|---|---|
| Selection_Method | | 1 = Switch Only<br>2 = Switch + Joystick<br>3 = Joystick + Switch<br>4 = Joystick Only<br>5 = Genie+<br>6 = Genie+ Joystick Only<br><br>Default: Genie+ |
| Time_Base | | 1 = 250ms,<br>2 = 300ms,<br>3 = 350ms,<br>4 = 450ms, (Default)<br>5 = 600ms,<br>6 = 800ms,<br>7 = 1.2sec,<br>8 = 2 sec |
| Avoid_Accidental_Hits | | 1 = No accidental hits are not avoided (default)<br>2 = Yes, during mode selection.<br>3 = Yes, during startup<br>4 = Yes, both<br><br>Default: No |
| Switch 2 | | Unused in Genie+ |
| Power_Down_Timer | | N = not required<br>Y = yes, power down the Genie+ after 20mins of no use.<br><br>Default: Yes |
| Exit_Time | | 1 = 500ms<br>2 = 1 sec exit time<br>3 = 2 sec (default)<br>4 = 3 sec exit time<br>5 = 4 sec exit time<br>6 = 5 sec exit time<br>7 = 6 sec exit time<br>8 = 7 sec exit time |

## Parameter Adjustment Table – Page Two

| Parameter (page 2) | Genie | Value and meaning |
|---|---|---|
| Alarm | | 1 = No Alarm<br>2 = Audible Alarm<br>3 = Audible Alarm + Text Message<br><br>Default: 1, no alarm |
| Max_Profile_Number | | Can be set from 1-5.<br><br>Default: 5 |
| RJM | | N = not fitted<br>Y = yes, RJM is fitted<br><br>Must be set to Y to enable the DX RJM module.<br><br>Default: Y |
| Enable_My_Set | | N = No, do not use my set of operation modes<br>Y = yes, use my set of operation modes (selection method must be either 1,2,3 or 4).<br><br>Default: No |
| Joystick_Threshold | | Unused in Genie+ |
| External_Device | | 1 = Standard connection.<br>2 = 2 cable connection to a DynaVox communication aid.<br>3 = 2 cable connection to a GEWA Progress ECU<br><br>Default: 1 |

## Parameter Adjustment Table – Page Three

| Parameter (page 2) | Genie | Value and meaning |
|---|---|---|
| Back Rest | | Can be set from 1-6. A setting between 1 and 5 selects the physical actuator associated with this symbol. Setting to 6 disables this symbol from the seat function display.<br><br>Default: 1 |
| Back Tilt | | Can be set from 1-6. A setting between 1 and 5 selects the physical actuator associated with this symbol. Setting to 6 disables this symbol from the seat function display.<br><br>Default: 2 |
| Seat Height | | Can be set from 1-6. A setting between 1 and 5 selects the physical actuator associated with this symbol. Setting to 6 disables this symbol from the seat function display.<br><br>Default: 5 |
| Foot Rest_Left | | Can be set from 1-6. A setting between 1 and 5 selects the physical actuator associated with this symbol. Setting to 6 disables this symbol from the seat function display.<br><br>Default: 3 |
| Foot Rest_Right | | Can be set from 1-6. A setting between 1 and 5 selects the physical actuator associated with this symbol. Setting to 6 disables this symbol from the seat function display.<br><br>Default: 4 |
| RJM_Mouse_Speed | | Sets the base mouse speed when using an RJM joystick.<br>1 = Slow<br>2 = Medium<br>3 = Fast<br><br>Default: 1 |

## Creating Your Own Set of Operation Modes

With the Genie+ it is possible to choose not only what operation modes will be accessible for a particular user but also in what order the will be presented.

The Genie+ program page 4 allows the technician to create a unique set of operation modes.

### To create your own set of modes

1. Select page 4 within the program mode (4 flashes of the decimal point)
2. The symbol in the 7 segment display indicates the current “first” mode.
3. Use short clicks of the Genie+ switch to display the “first” mode you want.
4. When it is displayed press and hold the Genie+ switch until the Genie+ beeps. The chosen mode for this step is stored and a new step begins.
5. Repeat from step 3 until you have entered all the steps of your set. Note that you can not exceed 6 steps, i.e. no more than 6 modes anyway.
6. When you have finished entering your set press and release the program switch. The steps are stored in the Genie+ and normal operation resumes.
7. To enable your set you must re-enter program mode and change the parameter ‘Enable_My_Set’ from ‘N’ to ‘Y’. The Selection_Method parameter must be either 1,2,3 or 4.

**Note**:

- When you store the first step of a new set the previous set is erased. In other-words you cannot edit a set, just replace it.

## Resetting Parameters to Factory Values

At some stage you may want to reset all the parameters to the factory values. The factory values are those listed as the default values in the parameter adjustment tables.

### To reset the parameters to factory values

1. Make sure the Genie+ is on.

2. Make sure the Genie+ is not in program mode.

3. Press and hold the program switch until the Genie+ beeps (about 10 seconds). Keep holding the program switch for a further 5 seconds until the Genie+ beeps a second time. Release the program switch.

4. All parameters have been reset to their factory (default) values.

Note: Be absolutely sure about resetting a user's parameter settings. If you have not written them down somewhere there is 'no way back'.

# 5 Programming the DX System

**Note !!**
The Genie+ is a third party product manufactured by Unique Perspectives Ltd. The Genie+ contains a DX User Control Module (UCM) whose programming is loosely based on the DX Dolphin remote. The following documentation applies to the Genie+ and has been referenced from the DX Dolphin Remote manual.

This section only deals with programming the DX system for optimum wheelchair performance. For programming of the Genie+ operation refer to chapter 4 in this section.

**Warning !!**
Incorrect or inappropriate programming of the DX System can put the wheelchair into a dangerous state. Unique Perspectives accept no responsibility or liability for accidents caused by incorrect programming. This section must be read and understood before attempting to program the DX System

Ensure that the programmed DX System complies with all prevailing regulatory requirements for your country and application

## Downloading the correct DX wizard file

**Note:** Normally it is not necessary to carry out this task as it will already have been done by your supplier.

Every master remote contains a DX wizard file that specifies all the data required to safely operate the wheelchair in question. It is not possible to use a handset (or Genie+) from one type of wheelchair on another without first downloading the correct wizard file. For example a handset fitted to a Cruiser Plus will not drive a Storm and visa versa.

When you order a Genie+ specify the type of wheelchair that you are connecting it to and the Genie+ will be provided with the correct file already downloaded.

If you have already purchased a Genie+ or wish to use one on another type of wheelchair contact technical@click2go.ie and we will email you the correct file if available.

If the correct file is not available there are 2 possible procedures for creating it and these are explained in the following pages.

# Procedure 1: Copying the file from the original handset into the Genie+ and editing it (Preferred).

### Upload the wizard file from the existing Handset

1. Connect the DX Wizard dongle into the printer port and connect the DX serial cable from the serial port of your computer and insert it into the programming socket on the Handset. Turn on the handset.

2. Open the DX wizard program and choose “Read Wheelchair” from the “Wheelchair” option in the menu bar.

### Edit the wizard file with Genie+ specific options

3. Double click on “UCM Remote” in the program modules list.

4. Ensure that the following 8 parameters are set according to the table below:

| Parameter | Value |
|---|---|
| Max Profile Number | 5 |
| Wrap Profiles | Yes |
| Change Prof Driving | Yes |
| Allow Non Driv Prof | Yes |
| Joystick Actuator | Yes |
| Joystick Source (RJM) (for every profile) | Yes |
| Sleep Enable | No |
| Lock Enable | No |

### Download the edited wizard file to the Genie

5. Replace the existing handset with the Genie+ and connect the DX serial cable to the programming socket underneath the unit.

6. Turn on the Genie+ and download the edited file by choosing “Write Wheelchair” from the “Wheelchair” option in the menu bar.

7. After download is complete turn off and on the Genie+ to complete the process and confirm that the Genie+ is driving the wheelchair and that items can be selected from the menu.

**NOTE 1:** You must have a Wizard program with OEM options in order to edit the above parameters. If you do not email your wizard file to technical@click2go.ie and we will do the changes for you.

**NOTE 2:** If the original handset has a Rev A or older UCM module you will be prompted to convert the file to a Rev C version when you try to download it to the Genie+. In this instance we recommend you use procedure 2 below or contact technical@click2go.ie and we will do the conversion for you.

## Procedure 2: Editing the file in the Genie+ to match the original handset

### Upload the wizard file from the existing Handset

1. Connect the DX Wizard dongle into the printer port and connect the DX serial cable from the serial port of your computer and insert it into the programming socket on the Handset. Turn on the handset.

2. Open the DX wizard program and choose “Read Wheelchair” from the “Wheelchair” option in the menu bar.

3. Print out the wizard file.

### Upload the wizard file from the Genie+ and edit it

4. Connect the DX serial cable from the serial port of your computer and insert it into the programming socket on the Genie+. Turn on the Genie+.

5. Open the DX wizard program and choose “Read Wheelchair” from the “Wheelchair” option in the menu bar.

6. Double click on “Power Module” in the program modules list. Edit all the parameters to match those on the print out.

7. Double click on “UCM Remote” in the program modules list. Edit all the parameters to match those on the print out except those listed in the table below which should already be set to the values shown.

| Parameter | Value |
|---|---|
| Max Profile Number | 5 |
| Wrap Profiles | Yes |
| Change Prof Driving | Yes |
| Allow Non Driv Prof | Yes |
| Joystick Actuator | Yes |
| Joystick Source (RJM) (for every profile) | Yes |
| Sleep Enable | No |
| Lock Enable | No |

8. Turn on the Genie+ and download the edited file by choosing “Write Wheelchair” from the “Wheelchair” option in the menu bar.

9. After download is complete turn off and on the Genie+ to complete the process and confirm that the Genie+ is driving the wheelchair and that items can be selected from the menu.

**NOTE 1:** You must have a Wizard program with OEM options in order to edit the above parameters. If you do not email your wizard file to technical@click2go.ie and we will do the changes for you.

# Programming Methods & Tools

## Dynamic Wizard

The Wizard is a PC based tool suited to programming production runs of identical wheelchairs or modules, or individual highly customised wheelchairs. The Wizard is available from Dynamic Controls in several versions:

| | |
|---|---|
| OEM | Generally used by the wheelchair manufacturer. Able to program a wide range of parameters |
| DEALER | Similar in function to above, but with a reduced range of programmable options. This ensures that options that the manufacturer wishes to keep control of cannot be disturbed. Parameters that may cause hazards or require special expertise to be set are not available to adjust. |
| ENHANCED DEALER | As above but with ability to edit parameters that relate directly to wheelchair accessories (e.g. actuators) |
| FACTORY | Can only replace Standard or Custom Wheelchair programs. No editing or diagnostics available. |

**Warning:** The Wizard is a very powerful tool and as such requires well trained operators and a disciplined approach to usage and distribution

The DX Wizard Manual should be read and understood before attempting to use it.

## Hand Held Programmer (HHP)

The DX Hand Held Programmer (HHP) is the normal programming tool used by dealers, allowing easy adjustment of all commonly adjusted Drive Program parameters.

**Warning:** The DX HHP is for use only by wheelchair manufacturers, their authorised dealers and support personnel. It is not for use by the wheelchair user.

The DX HHP manual should be read and understood before attempting to use it.

## Auto Download – replacing other modules

The DX System has a feature called Auto Download. It is designed to minimise the programming requirements associated with Module servicing by down loading the correct programming to a replacement DX Module.

Auto Download is achieved by DX remotes containing both their own programming and also a backup copy of the programmed data for all other DX Modules in the system. When a module is swapped, or a check sum error found in a module, the DX Remote automatically down loads its backup copy to the module. The Auto down load occurs immediately on power up after the module has been replaced. This applies to all DX modules **except** a DX master remote. I.e. When you replace a Genie+ it will need re-programmimg.

**Warning!!**
When a Genie+ is replaced it will perform an Auto Down load to all DX modules. This may result in incorrect and dangerous programming for a particular wheelchair system if the wheelchair program installed in the Genie+ is not suitable for that wheelchair system.

Do not attempt to drive or test the DX system before the correct and suitable wheelchair program has been installed in the Genie+ using the Wizard.

See page 63 “Downloading the correct DX wizard file”.

After replacing any DX Module, turn the DX System off, then on again, to initiate the Auto down load of the DX Remote backup data. When a Auto Down load has occurred the status LED of the replaced module will flash. Turn the system off and them on again to clear this fault and complete the Auto Down load.

# 6 Battery Warnings, Diagnostics and Fault Finding

## Battery Condition Warnings

A battery warning is shown by the battery gauge flashing its LEDs.

### Battery high warning condition

This condition occurs when the battery voltage exceeds 28V, as measured by the PM. The cause can be:

- The wheelchair is still on charge and the batteries are full or faulty
- The batteries are overcharged
- The wheelchair is travelling down a slope and the batteries are full or faulty

The wheelchair will drive during this fault condition which will reset automatically when the battery voltage drops below 28v.

### Battery Low warning condition

This condition occurs when the battery voltage drops below 23.3V, when the wheelchair is still.

The cause can be:

- If the battery gauge flashes with orange or green LEDs lit, but the cause is not due to a battery high fault, the battery or battery wiring may be faulty.
- If the battery gauge flashes with just the 1 or 2 red LEDs lit after stopping the chair, the battery may be too small for the wheelchair type, or the battery may be old or damaged.

A battery low warning normally coincides with a low capacity warning.

### Low capacity warning condition

When the calculated available battery capacity drops below 10% of full capacity the bottom red LED flashes.

The wheelchair will drive during this fault condition but it shows that the battery is in the reserve capacity range and battery capacity will begin to reduce very

rapidly. The low capacity warning will not stop until the batteries have been recharged adequately.

## Diagnostics and Fault Finding

DX System diagnostics can be examined from two platforms : from the Flash Codes signaled with the system status LED on the Genie+ (and on the HHP and from the Wizard which can provide more detailed information about the fault.

### Flash Code

Any fault condition on the DX system will cause the Genie+'s system status LED to flash. Flashing occurs in burst of flashes separated by a two second pause. The number of flashes in each burst is referred to as the Flash Code and indicates the nature of the fault. The title of the Flash Code fault is also displayed by the HHP if connected to the faulty wheelchair.

Faults that affect the safety of the chair will cause the chair to stop while less critical ones will be indicated but allow the chair to continue driving. Some faults will automatically clear when the fault condition is removed, in which case the System Status LED will extinguish. Other faults are latched and must be cleared by turning the DX System off, waiting for two seconds, and then turning it back on again.

| DX System Status LED Flash Code | Likely cause of condition and possible action |
|---|---|
| 1 | **DX Module Fault (see limp Mode below)**<br>Cause An auto download has occurred<br>Action ✍ Turn the Genie+ off then on again.<br>Cause The DX System is not correctly programmed.<br>Action ✍ Try reprogramming the DX System<br>Cause Connection between DX Modules may be faulty, or there may be an internal fault in a Module.<br>Action ✍ Check DXBUS connections and replace if necessary<br>✍ If the Status LED on another Module is flashing, replace the Module<br>✍ An expected module may not be present (e.g. the Actuator Module) |
| 2 | **DX Accessory Fault** |

Cause There is a fault in an accessory device attached to a DX Module (excluding the PM). Examples of faults in accessory devices may be: the clutch is, or has been, disengaged; a light bulb is short or open circuit; an actuator terminal is shorted to Battery +

Action
- ✍ Check all accessory devices connected to your DX System

3 **Left (M1) Motor Fault**

Cause The connection from the PM Left (M1) connector to its associated motor, or the motor itself, is defective. The connection is either open or short circuit.

Action
- ✍ Disconnect the left motor plug and check continuity between the motor pins on M1

4 **Right (M2) Motor Fault**

Cause The connection from the PM Right (M2) connector to its associated motor, or the motor itself, is defective. The connection is either open or short circuit.

Action
- ✍ Disconnect the right motor plug and check continuity between the motor pins on M2

5 **Left (M1) Park Brake Fault**

Cause The M1 plug connection to its associated park brake is either open or short circuit

Action
- ✍ Disconnect the M1 plug and check continuity between the two positronic park brake pins

6 **Right (M2) Park Brake Fault**

Cause The M2 plug connection to its associated park brake is either open or short circuit

Action
- ✍ Disconnect the M2 plug and check continuity between the two positronic park brake pins

7 **Low Battery Fault**

Cause The battery charge is not sufficient to allow safe driving. It has fallen below 17V

Action
- ✍ Check the battery connection and terminals. The battery voltage should be similar when the battery is on charge, and when it isn't.
- ✍ Check that fuses have not blown, or circuit breakers tripped.
- ✍ Replace battery if worn out or if capacity is insufficient for the user's needs.

**Note:** The wheelchair will behave sluggishly and the Battery Gauge will flash indicating low battery voltage prior to this fault.

8 **Overvoltage Fault**

Cause The battery voltage has exceeded 32V

Action
- ✍ If this fault occurs during battery charging, the battery charger is defective or incorrectly adjusted.
- ✍ Check the battery chargers open circuit voltage is in accordance with the battery manufacturers limits, and is less than 32V

Cause The battery connector is making intermittent contact when the wheelchair is stopped, or travelling down a slope.

Action
- ✍ Check that the battery wiring and terminating is secure.

9 **CANL Fault (see limp mode)**

Cause
1. An invalid voltage has been detected on the DXBUS CANL line.
2. Communication is not possible using the CANL wire.

Action
- ✍ Check the continuity of the DXBUS cable
- ✍ Check for shorts between the DXBUS pins. An open or short circuit on another DX Module can cause this fault.

10 **CANH Fault (see limp mode)**

Cause
1. An invalid voltage has been detected on the DXBUS CANH line.
2. Communication is not possible using the CANH wire, or the CANJ and CANL wires are shorted together.
3. Hazard lights were turned on when the DX system was turned on.
4. The CANH is used to generate a Kill signal by any DX Module which detects and unsafe condition, or by an external device such as an emergency stop switch. The CANH wire is pulled to either Battery + or Battery – and causes the DX System to shut down.

Action
- ✍ Check continuity of the DXBUS cable.
- ✍ Check for shorts between the DXBUS pins. An open or short circuit on another DX Modules can cause this fault.
- ✍ If the hazard lights were already switched on when the DX system was turned on, Flash code 10 and limp Mode may result. To clear this fault, turn the Hazard lights off, then turn the DX System off then on again.
- ✍ If generated by a Kill signal, the cause of the fault is severe.

11 **Stall Timeout Fault**

Cause: The motor current has been at, or close to, current limit for longer than the Stall Timeout parameter value.

Action
- ✍ Turn the DX System off then on again.

12 **Module Mismatch**

Cause: There is a compatibility problem between DX Modules in the system. The wheelchair will be disabled.

Action
- ✍ Consult your Dynamic service center.

Cause: The data held by the DX UCM for another DX Module is corrupt or incompatible with that module

Action
- ✍ Reprogramming the wheelchair system may correct this problem.

## Limp Mode

If the DX System detects some faults, it will revert to Limp Mode. This is a reduced speed mode which recognises problems, but allows the wheelchair user to limp home, where the problem can be assessed.

# 7 Maintenance

1. The Genie+ system should be regularly checked for integrity. Loose, damaged or corroded connectors or terminals, or damaged cabling should be reported to your Service Centre and be replaced immediately.

2. The cabling of Genie+ System including Switch cables, DXBUS cable, Charger cable and Battery cables, should be regularly checked for integrity. They should never be loose. Cables should be neatly attached to the wheelchair frame and mounts so that no possibility exists for a cable to become snagged on the moving parts of the wheelchair itself, the person sitting in the wheelchair, and/or items external to the wheelchair such as door handles etc.

3. All switches connected to the Genie+ should be regularly tested to ensure that they function correctly.

4. During storage and transport of your wheelchair ensure that there is no possibility that The Genie+ switch can inadvertently by pressed thereby causing the chair to turn on and possible enter a drive state. Always disengage the motor gears and disconnect any external switches.

5. Under no condition should a latching switch be connected to the Genie+. Only connect non-latching switches.

6. The Genie+ components and other wheelchair parts should be kept free of dust, dirt and liquids. If necessary wipe with a cloth dampened with warm water or alcohol. **Do not** use solvents or abrasive cleaners.

7. Where any doubt exists, consult your nearest Service Centre or Agent.

8. There are no user-serviceable parts within the Genie+. Do not attempt to open the case.

9. In accordance with the requirements of CE marking of this device and the Company's policy, it is requested that re-occurring faults or defects are reported back to Unique Perspectives Ltd.

---

**Warning !!** If the Genie+ is damaged in any way, or if internal damage may have occurred (for example by being dropped), have it checked by qualified personnel before operating.

---

# 8 Safety and Misuse Warnings

Do not install, maintain or operate this equipment without reading, understanding and following the proper instructions and manuals, otherwise injury or damage may result.

The completed installation must be thoroughly checked, and all programmable options must be correctly adjusted for safe operation prior to use.

A warning must be conveyed to the wheelchair operator that the controller could cause the chair to come to a sudden stop. In situations where this may affect the safety of the user, this will require the fitting and wearing of a seat belt.

Performance adjustments should only be made by professionals of the health care field or persons fully conversant with this process and the driver's capabilities. Incorrect settings could cause injury to the driver, bystanders, damage to the chair and surrounding property.

After the wheelchair has been set up, check to make sure that the wheelchair performs to the specifications entered in the programming procedures. If the wheelchair does not perform to specifications, turn the wheelchair off immediately and re-program. Repeat procedure until the wheelchair performs to the specifications.

Do not operate the DX system if it behaves erratically, or shows abnormal response, heating, smoke or arcing. Turn the system off, disconnect the battery or open the battery overload switch, and consult your service agent.

Do not operate your DX system if the battery is nearly flat as a dangerous situation may result due to a loss of power in an inopportune place.

Ensure the controller is turned off when not in use.

No connector pins should be touched, as contamination or damage due to electrostatic discharge may result. Dummy sockets in unused DXBUS connectors should be left in place unless a new module is added to the system.

Whilst designed to resist water penetration, under certain conditions moisture might enter the Genie+. Any spillage's over the Genie+ should be wiped dry without delay. The Genie+ may be used outdoors in light drizzle conditions but should be protected from rain.

Most electronic equipment is influenced by Radio Frequency Interference (RFI). Caution should be exercised with regard to the use of portable communications equipment in the area around such equipment. While the manufacturer has

made every effort to ensure that RFI does not cause problems, very strong signals could still cause a problem. If RFI causes erratic behavior, shut the wheelchair off immediately. Leave off while transmission is in progress.

In the event of a fault indicator flashing while driving (battery gauge and/or status LED), the user must ensure that the system is behaving normally. If not, the system must be turned off and a Service Agent called immediately.

Report any malfunctions immediately to your Service Agent.

## Know the risks and limitations

Like any mechanical propelled vehicle there are certain risks involved.

The driver is responsible for any damage or injury that may occur to a party as a result of using a powered wheelchair. If the driver cannot assume responsibility due to age or disability then a carer must be present and be able to take over control either using a stop switch or a dual control in case of an emergency. You may wish to consider taking out insurance to cover any claims arising from such an incident.

The most sensitive part of a Genie+ system is the joystick element itself. The owner or carer must assume responsibility for regularly checking the integrity of the joystick element and report any problems to the service agent immediately.

# 9 EC Declaration Of Conformity

## Rationale

The Unique Perspectives Ltd. Genie+ device is designed for the intended purpose of providing an alternative method of controlling the function of a powered wheelchair. A wheelchair is considered a medical device as defined within Council Directive 93/42/EEC concerning medical devices.

The Genie+ is designed for use in combination with a Controls Dynamic DX control system, incorporates a DX UCM and uses a standard DXBUS connector.

The Genie+ replaces the standard Controls Dynamic's Joystick.

As the intended purpose is to control a powered wheelchair, and as the Genie+ meets the definition of an "Accessory" of a medical device, as defined at Article 1 paragraph 2 b) of the Directive, the Genie+ has been CE marked as a medical device in accord with the Directive requirements.

## Classification

The Unique Perspectives Ltd. Genie+ control system is intended to provide an alternative operating / control system for those who would find it difficult, painful, or impossible, due to an injury or handicap, to operate a standard powered wheelchair control. It is intended to be used in combination with Class I medical devices only.

The Genie+ is classified as a Class I device.

## Compliance Certificate

The compliance certificate issued by Unique Perspectives Ltd. does not relieve a wheelchair manufacturer from compliance testing their particular wheelchairs.

A manufacturers wheelchair must meet the requirements of the directive before fitting of the Unique Perspectives Ltd. Genie+.

As the Genie+ is an 'Accessory' to a medical device re-testing of a manufacturer's wheelchair with the Genie+ fitted is not a requirement.

Council Directive 93/42/EEC of 14 June 1993
concerning medical devices.

Unique Perspectives Limited, Ballyclovan Cottage, Ballyline, Callan, Co. Kilkenny, Ireland, declare that:

? the Unique Perspectives Limited Genie+ device, as described within the technical file TF.1, conforms to Class I requirements of Council Directive 93/42/EEC of 14 June 1993 concerning medical devices;

? is in conformity with the provisions of that Directive and, where such is the case, with the national standards transposing harmonized standards as noted within the technical file TF 1;

? the requirements of Annex I, Annex VII, Annex IX have been followed for Class I devices and registration requirements of Article 14 have been notified to the Rep. of Ireland Competent Authority;

? the Unique Perspectives Limited Genie+ device is an Accessory of a medical device;

? the Declaration covers all Unique Perspectives Limited Genie+ devices placed on the market on or after May 2006 and until such time as a renewed conformity declaration is raised.

This declaration is signed on behalf of Unique Perspectives Limited by:

Signed _____

Name _____

Company Position _____

Date May 2006 _____

# 10 Warranty

All equipment supplied by Unique Perspectives Ltd is warranted by the company to be free from faulty materials or workmanship. If any defect is found within the warranty period of 6 months, the company will repair the equipment, or at its discretion, replace the equipment without charge for materials and labor.

The warranty is subject to the conditions that the equipment:

- Has been correctly installed
- Has been used solely in accordance with this manual.
- Has been properly connected to a DX Power Module in accordance with this manual.
- Has not been subjected to misuse or accident, or been modified or repaired by any person other than someone authorised by Unique Perspectives Ltd.
- Has been used solely for the driving of electrically powered wheelchairs in accordance with the wheelchair manufacturer's recommendations.

# 11 Sales and Service Information

For Sales and Service advice, or in case of any difficulty, please contact:

Unique Perspectives Ltd.
Ballyclovan
Callan
Kilkenny
Ireland

Telephone: +353 56 7725913
Fax: +353 56 7725936

www.click2go.ie
info@click2go.ie

Note: The controller should be clearly labeled with the manufacturer's service agent's telephone number.

# 12 Genie+ Quick Reference Guide

| Page 1 Parameters | Value and meaning |
|---|---|
| | 1 = Switch Only<br>2 = Switch + Joystick<br>3 = Joystick + Switch<br>4 = Joystick Only<br>5 = Genie+<br>6 = Genie+ Joystick Only<br>Default: 5 Genie+ |
| | 1 = 250ms,<br>2 = 300ms,<br>3 = 350ms,<br>4 = 450ms, (Default)<br>5 = 600ms,<br>6 = 800ms,<br>7 = 1.2sec,<br>8 = 2 sec |
| | 1 = No accidental hits are not avoided (default)<br>2 = Yes, during mode selection.<br>3 = Yes, during startup<br>4 = Yes, both |
| | Not used in Genie+ |
| | N = not required<br>Y = yes, power down the Genie+ after 20mins of no use.<br><br>Default: Yes |
| | 1 = 500ms<br>2 = 1 sec exit time<br>3 = 2 sec (default)<br>4 = 3 sec exit time<br>5 = 4 sec exit time<br>6 = 5 sec exit time<br>7 = 6 sec exit time<br>8 = 7 sec exit time |

| Page 2 Parameters | Value and meaning |
|---|---|
| | 1 = No Alarm<br>2 = Audible Alarm<br>3 = Audible Alarm + Text Message<br><br>Default: 1, no alarm |
| | Can be set from 1-5.<br><br>Default: 5 |
| | N = not fitted<br>Y = yes, RJM is fitted<br><br>Must be set to yes to enable the DX RJM module.<br><br>Default: Yes. |
| | N = No, do not use my set of operation modes<br>Y = yes, use my set of operation modes.<br><br>Default: No |
| | Not used in Genie+ |
| | 1 = Standard connection.<br>2 = 2 cable connection to a DynaVox communication aid.<br>3 = 2 cable connection to a GEWA Progress ECU<br><br>Default: 1 |

**To set a parameter**

1. Make sure the Genie+ is on.
2. Select program mode by holding the program switch for 10 seconds until you hear a beep.
3. Page 1 is automatically selected.
4. The number of times the decimal point flashes indicates what page you are on. To select the next page press the program switch.
5. Select the parameter by clicking the Genie+ switch until the corresponding number of battery LEDs are lit.
6. Change the parameter by holding down the Genie+ switch until the unit beeps and the value displayed in the 7 segment display changes.
7. To exit program mode step through the remaining page(s) by clicking the program switch until the decimal point flashes very rapidly.

**To reset the parameters to factory values**

1. Make sure the Genie+ is on.
2. Press and hold the program switch for 10 seconds until you hear a beep. Continue holding for a further 5 seconds until you hear a second beep.
3. Release the program switch and the decimal point flashes rapidly.
4. All parameters have been reset to their factory (default) values.

Before you do any programming, look at the these tables of parameters and know beforehand what changes you want to make. Be absolutely sure about changing a user's parameter settings. It is a good idea to photocopy this sheet and make note of a user's parameter settings.

**To modify the seat function list**

1. Enter program mode and select page 3.
2. The first five parameters determine what actuator is controlled by what seat symbol.
3. Change a parameter in the normal way (see opposite text). Each parameter can be set from 1 to 6. Setting a parameter to 6 means that seat symbol will not be used.
4. See text for further explanation.

**To create your own set of operation modes**

5. Enter program mode and select page 4.
6. The symbol in the 7 segment display indicates the first mode.
7. Use short clicks of the Genie+ switch to select the mode you want for this step.
8. When it is displayed press the Genie+ switch for a long time, until the Genie+ beeps. The chosen mode for this step is stored and a new step begins.
9. Repeat from step 3 until you have entered all the steps of your set. Note that you can not exceed 6 steps.
10. When you have finished entering your set press and release the program switch. The steps are stored in the ClickToGo and normal operation resumes.
11. To enable your set you must re-enter program mode and change the parameter 'Enable_Set' on page 2 from 'N' to 'Y'. You must also set the Selection_Method parameter to 1,2,3 or 4.

# IMPROVEMENTS TO GENIE+

Nov 07

Genie+ and Genie upgraded to DX UCM2. This means that from Units with serial Number 100 and higher the Genie can be used with the DXGB power module, i.e. Gearless brushless motors.

Nov 07
Genie+ Joystick only mode improved so that switch has no effect except to switch on. Previously the switch worked as a selection as well as a backwards flick. This was seen to cause problems for those wishing to use a joystick with a push for switching on but using a flick for mode selection.